

## Diaphragm Seal Options

		SS Armored Capillary	SS Armored Capillary w/PVC Sleeve	Pipe Plug for Flushing Conn.	Top Housing 316 SS	Top Housing Monel	Top Housing Hastelloy C	SS Clamp/Flange Rings Bolts Nuts	High Pressure Clamp Rings	Clamping Bolts 300 Series SS	Cleaning For Gaseous Oxygen or Strong Oxidizing Agents <sup>(4)</sup>	Inst. Welded to Seal	Positive Material Ident. (PMI)
	CODE	1115A	1115P	PU	YT	YM	HB	SE	HP	SB	6B	DU	MQ
SEAL TYPE	TYPE NUMBER												
THREADED	100	•	•		•			•	•(1)	•	•	•	•
	101	•	•	•	•			•	•(1)	•	•	•	•
	200	•	•		•	•		•	•(1)	•	•	•	•
	201	•	•	•	•	•		•	•(1)	•	•	•	•
	300	•	•		•			•	•(1)	•	•	•	•
	301	•	•	•	•			•	•(1)	•	•	•	•
	104	•	•		•							•	•
	310	•	•			STD						•	STD <sup>(3)</sup>
	315	•	•			STD						•	STD <sup>(3)</sup>
	311	•	•			STD						•	•
	312	•	•			STD						•	•
	330	•	•									•	•
	400	•	•			STD		•	•	•(2)	•	•	•
	401	•	•	•		STD		•	•	•(1)	•	•	•
	500	•	•			STD		•				•	•
	501	•	•	•		STD		•				•	•
	510	•	•			STD				•(1)		•	•
511	•	•			STD						•	•	
740	•	•			STD						•	•	
741	•	•	•		STD						•	•	
FLANGED	102	•	•		•			•		•	•	•	•
	103	•	•	•	•			•		•	•	•	•
	202	•	•		•			•		•	•	•	•
	203	•	•	•	•			•		•	•	•	•
	302	•	•		•			•		•	•	•	•
	303	•	•	•	•			•		•	•	•	•
	106	•	•		•			•				•	•
	206	•	•		•			•				•	•
	402	•	•			STD		•				•	•
	403	•	•	•		STD		•				•	•
702	•	•			STD						•	•	
703	•	•	•		STD						•	•	
IN-LINE	105	•	•		•			•			•	•	•
	107	•	•		•						•	•	•
	108	•	•		•			•			•	•	•
	205	•	•		•			•			•	•	•
	207	•	•		•						•	•	•
	208	•	•		•						•	•	•
QUICK CONN.	320	•	•								•	•	•

**HOW TO ORDER:** (Refer To Table On Pages 2 & 3)

Typical ordering code: 10-102-SS-04T-XCGYT-150-RF

- 2 - From process connection size (code 10 = 1")
- 1 - From type number (code 102 = flanged seal with threaded diaphragm)
- 9 - From lower housing and diaphragm material (1st S = 316 st. st. diaphragm, 2nd S = 316 st. st. lower housing)
- 3 - From instrument connection size (code 04T = 1/2 NPT)
- 10 - Diaphragm seal assembly fill fluid & options; precede option code with 'X' (code CG from page 3 = glycerin instrument and seal fill, code YT from page 4 above = top housing 316 st. st.)
- 5 - If a flanged seal, select flange rating (code 150 = 150 ANSI B16.5 class flange)
- 6 - If for a flanged seal, select flange type (code RF = ANSI B16.5 raised face flange)

**NOTES:**

- (1) 5000 psi pressure rating
- (2) 9000 psi pressure rating
- (3) Standard with monel diaphragm
- (4) Fill must be XCF (Halocarbon)

## Ten Steps to Select a Diaphragm Seal



150, 300, 600, 900, 1500  
Metal diaphragm (750°) Teflon (-40/400°F)  
Kalrez (30/212°F) Viton (-40/350°F)



ASHCROFT OFFERS A WIDE SELECTION OF OPTIONS, SEE PAGE 4.

- 1 SEAL TYPE**  
When determining a seal type, two requirements must be considered:  
1. **Process Connection Type** –  
A **threaded design seal** connects directly to the process with a female or male NPT connection.  
A **flanged design seal** is attached to the process with a flange as specified in ASME B 16.5.  
An **in-line welded design seal** is suitable for flow-thru applications.  
2. **Type Number** –
- 2 PROCESS CONNECTION SIZE & TYPE**  
Select process connection size. If the requirement is for a threaded seal determine if a male or female connection is required.
- 3 INSTRUMENT CONNECTION SIZE**  
Determine if a 1/4 NPT or 1/2 NPT is required
- 4 DIAPHRAGM TYPE**  
Is the requirement for the diaphragm configuration threaded, welded, bonded or clamped to the top housing. Design types are:  
**Threaded Design:** ensures a positive sealing surface. The diaphragm can be replaced if damaged.  
**Welded or Bonded Design:** Metallic Diaphragm- welded to top housing.  
Elastomeric Diaphragm: bonded to top housing. Both ensure maximum leak integrity.  
**Clamped Design:** available with elastomeric diaphragms only. Diaphragm is clamped between top and bottom housing.
- 5 MAXIMUM ALLOWABLE WORKING PRESSURE**  
Ensure the instrument full scale range does not exceed the rated pressure of the diaphragm seal. Flanged seal class ratings are in accordance with ASME B16.5. If the diaphragm seal will be used with a differential pressure instrument, the instrument static pressure should not exceed the rated pressure of the diaphragm seal.
- 6 FLANGE TYPE**  
If the requirement is for a flanged seal, determine if a raised face, flat face or ring joint flange is required
- 7 PROCESS TEMPERATURE LIMITS**  
When selecting the required system assembly fill fluid, lower housing and diaphragm material, ensure minimum and maximum temperature limits are compatible with the selected fill fluid, diaphragm and lower housing materials. When the requirement is for a flanged seal, refer to ASME B16.5 for pressure and temperature limits.
- 8 TOP HOUSING, FLANGE OR CLAMP RING MATERIAL, CLAMP BOLTS**  
Consider environmental compatibility when selecting.  
**Flanged Type Seals:** Standard flange material is nickel plated carbon steel.  
**Threaded Type Seals:** When a clamped ring is offered, standard material is black painted carbon steel.  
**Clamping Bolts:** Standard bolt material is zinc plated carbon steel.  
See Diaphragm Seal Options page 4 when the standard material is not compatible with environmental conditions.
- 9 LOWER HOUSING & DIAPHRAGM MATERIAL**  
Both the diaphragm and lower housing are 'wetted parts' and must be compatible with the process media. Refer to our Corrosion Data Guide under Technical Info at [www.ashcroft.com](http://www.ashcroft.com) for material compatibility. Process temperature limits and concentration are a requirement when selecting lower housing and diaphragm material.
- 10 FILL FLUID**  
Consider process temperature and process media compatibility when selecting the system fill fluid. A fill fluid other than glycerin is required for vacuum and compound gauge ranges. Glycerin fill is not available when capillary is required between the instrument and diaphragm seal.

All specifications are subject to change without notice.  
All sales subject to standard terms and conditions.  
© Ashcroft Inc. 2015 04/15

Ashcroft Inc., 250 East Main Street, Stratford, CT 06614 USA  
Tel: 203-378-8281 • Fax: 203-385-0408  
email: [info@ashcroft.com](mailto:info@ashcroft.com) • [www.ashcroft.com](http://www.ashcroft.com)

